

**MERRITT COLLEGE SCIENCE & ALLIED HEALTH BUILDING SCIENCE EQUIPMENT &  
APPLIANCES**

**BID NO. 14-15/32**

**Peralta Community College District**

**333 East 8th Street, Oakland, CA 94606**

April 17, 2015

**ADDENDUM No. 3**

This addendum supersedes items of the original contract documents wherein it is inconsistent with it. All other conditions remain unchanged. The following changes, modifications, corrections, additions or clarifications shall apply to the contract documents and shall be made a part of and subject to all of the requirements thereof as if originally specified or shown. It is the responsibility of the bidder to review the list of attachments to ensure that the addendum is full and complete. This Addendum modifies the original Bid Documents for the above Bid. **Acknowledge receipt of this addendum in the space provided on the BID FORM. Failure to do so may subject Bidder to disqualification.**

**Bid document clarification**

1. Attachment 11 – Bid Elements, Item 13, Freezer LN2 (w/ back up dewar) Model 7452 has a variety of sizes for the back-up dewar, not specified in the original documents. **The preferred back-up dewar size is 10 Liters.**

**END OF ADDENDUM THREE**